



Food and Drug Branch – Device Recalls

CALIFORNIA DEVICE RECALL INFORMATION SHEET

Microbiologics Inc Recalls Enteric Pathogens Control Panel For Contamination Of Negative Control

Recall Date	Product Description	Recalling Firm	Recall Reason
2/21/2024	Microbiologics Enteric Pathogens Control Panel (Inactivated Pellet) REF 8184	MICROBIOLOGICS INC Saint Cloud, Minnesota	The negative control was contaminated with one of the positive control pathogens (Norovirus). This would result in users getting a positive call for Norovirus when running the negative control.

Recall Class	Product Identification	Distribution	Affected Dates
II	Microbiologics Enteric Pathogens Control Panel (Inactivated Pellet) Lot #8184-68, Exp. 5/31/2025, UDI-DI 70845357041448. UPDATE: The recall was expanded on 1/23/2024 to include the following additional lot numbers: 8184-65, exp. 3/31/2025; 8184-66, exp. 4/30/2025; 8184-67, exp. 5/31/2025; 8184-69, exp. 6/30/2025; 8184-70, exp. 8/31/2025; and 8184-71, exp. 8/31/2025.	64 Units in California	Expires March through August, 2025

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [FDA WEBSITE](#)

