



## California Device Recall Information Sheet

**Zimmer Surgical Inc Recalls Skin Graft Carrier for Incorrect Labeling of the Ridge Pattern Ratio.**

Recall Date	Product Description	Recalling Firm	Recall Reason
2/21/2024	<b>Skin Graft Carrier</b> 3:1 Dermacarrier, Model Number 00219501300	<b>Zimmer Surgical Inc</b> Dover, Ohio	This product is being recalled due to incorrect labeling. The product is labeled as 3:1 ratio, but the ridge pattern is a 1.5:1 ratio. The problem may lead to minor tissue damage or additional unplanned graft

Recall Class	Product Identification	Distribution	Affected Dates
II	<b>Skin Graft Carrier</b> UDI-DI 00889024378780 Lots 65292843 (Expires 16-Oct-2026) and 65390419 (Expires 18-Jan-2027)	6 Units in California California	February, 2024 and prior

For additional information, please visit the [FDA website](https://www.fda.gov)

