



Food and Drug Branch – Device Recalls

California Device Recall Information Sheet

Baxter Healthcare Recalls Live2000 Ventilator

Recall Date	Product Description	Recalling Firm	Recall Reason
7/3/2024	Live2000 Ventilator Product Code: MS-01-0118, which is contained in the Life2000 Ventilator System, with Product Codes BT-20-0002, BT-20-0002A and BT-20-0002AP. The Life2000 ventilation system includes the Life2000 ventilator which has its individual label.	Baxter Healthcare Corporation Deerfield, Illinois	Potential for certain Life2000 ventilator systems to either fail to charge or have intermittent charging behavior due to damage to the battery charger dongle. Damage of the battery charger dongle prevents the ventilator's internal battery from charging.

Recall Class	Product Identification	Distribution	Affected Dates
I	Live2000 Ventilator Live2000 Ventilator: UDI Number: 00887761978089; Product Code: MS-01-0118. Live2000 Ventilator System: a. Product Code: BT-20-0002; Serial Numbers: 123300001188	2510 Units Nationwide	June 2024 and prior.

For additional information, please visit the [FDA website](#)

CDPH Food and Drug Branch
 MS 7602 • P.O. Box 997435 • Sacramento, CA 95899-7435
 (916) 650-6500 • (916) 650-6650 FAX
 Internet Address: www.cdph.ca.gov

