



[Food and Drug Branch – Device Recalls](#)

California Device Recall Information Sheet

Covidien Recalls Mcgrath Mac Video Laryngoscope

Recall Date	Product Description	Recalling Firm	Recall Reason
8/28/2024	Mcgrath Mac Video Laryngoscope McGrath Mac Video Laryngoscope, REF: 300-000-000, Non-Sterile, Rx Only	COVIDIEN Boulder, CO	Battery management system within Laryngoscope devices may deplete below the design threshold which may result in thermal event followed by risk of explosion.

Recall Class	Product Identification	Distribution	Affected Dates
I	Mcgrath Mac Video Laryngoscope UDI: 15060272980020 05060272980023 F0602729800201	75,544 units Units Nationwide including C	August 2024 and Prior

For additional information, please visit the [FDA website](#).

