



Food and Drug Branch – Device Recalls

California Device Recall Information Sheet

Karl Storz Endoscopy Recalls Optical Scissor And Grasping Forceps Due To Issue With Evidence Of Validation

Recall Date	Product Description	Recalling Firm	Recall Reason
4/10/2024	Karl Storz Se & Co. Kg, Optical Scissor, Ce REF 115400S	KARL STORZ ENDOSCOPY El Segundo, California	Inadequate reprocessing validation evidence.
4/10/2024	Karl Storz Se & Co. Kg, Grasping Forceps REF 11003MB, Flexible 1 mm, Rx Only	KARL STORZ ENDOSCOPY El Segundo, California	Inadequate reprocessing validation evidence.

Recall Class	Product Identification	Distribution	Affected Dates
II	Karl Storz Se & Co. Kg, Optical Scissor, Ce UDI: 04048551259788/ All Lots	0 Units in California	February, 2024 and prior
II	Karl Storz Se & Co. Kg, Grasping Forceps UDI: 04048551050255/ All Lots	17 Units in California	February, 2024 and prior

For additional information, please visit the FDA website [Optical Scissors](#), [Grasping Forceps](#)

