



[Food and Drug Branch – Device Recalls](#)

California Device Recall Information Sheet

Medline Industries, Lp - Northfield Recalls Medical Convenience Kits

Recall Date	Product Description	Recalling Firm	Recall Reason
8/14/2024	Medical Convenience Kits Medline procedure kits labeled as: 1) NEPHROLOGY PACK, Pack Number DYNJ59089; 2) NERVE BLOCK TRAY, Pack Number DYNJRA1221A	MEDLINE INDUSTRIES, LP - NORTHFIELD Northfield, IL	Medical convenience kits contain plastic syringes affected by the FDA Safety Alert issued on March 19, 2024. Leaks, breakage, and/ other quality issues have been identified that may pose a risk to patient health.

Recall Class	Product Identification	Distribution	Affected Dates
I	Medical Convenience Kits DYNJ59089 , Lot Number 23IBF074 ; DYNJRA1221A , Lot Number 20IDB255 ; DYNJRA1221A , Lot Number 22BDA978 ; DYNJRA1221A , Lot Number 22KDA950 ; DYNJRA1221A , Lot Number 22LDA146 ; DYNJRA1221A , Lot Number 23ADB422 ; DYNJRA1221A , Lot Number 23BDB453 ; DYNJRA1221A , Lot Number 23IDA424 ; DYNJRA1221A , Lot Number 23JDB010	1,679,067 Units Nationwide including C	August 2024 and Prior

For additional information, please visit the [FDA website](#).

