



[Food and Drug Branch – Device Recalls](#)

California Device Recall Information Sheet

Medtronic Minimed, Inc. Recalls Minimed 620g Insulin Pump

Recall Date	Product Description	Recalling Firm	Recall Reason
10/9/2024	<p>Minimed 620G Insulin Pump MiniMed 620G Insulin Pump, REF: MMT-1510, MMT-1710, MMT-1750; MiniMed 630G Insulin Pump, REF: MMT-1515, MMT-1714, MMT-1715, MMT-1754, MMT-1755; MiniMed 640G Insulin Pump, REF: MMT-1511, MMT-1711, MMT-1512, MMT-1712, MMT-1751, MMT-1752; MiniMed 720G Ins</p>	<p>MEDTRONIC MINIMED, INC. Northridge, CA</p>	<p>Insulin pumps that have been dropped, bumped, or experienced physical impact may have damage to internal electrical components, which may cause reduced pump battery life. Since the defect reduces the battery life overall, it shortens the time between low battery alarms and increases the frequency with which the user must insert a new battery. Further, the defect may also deliver the low battery alarm when there is significantly less time before the battery completely runs out, from the expected up to 10-hour buffer the low alert is supposed to signal, leading to device powering down and stopping insulin delivery unexpectedly. Per the firm, a subset of pumps with the</p>

			defect showed the low alert alarm was delivered ~2.5 hours before the device shut off, however, patient reports indicate the time could be even shorter. Replacing the battery will not resolve this issue and the short battery life and the truncated low battery alert lead time will continue to occur. Early battery depletion could result in power loss and insulin delivery could be interrupted, resulting in under delivery of insulin potentially leading to hyperglycemia and/or DKA.
--	--	--	---

Recall Class	Product Identification	Distribution	Affected Dates
I	Minimed 620G Insulin Pump REF/UDI-DI(GTIN):MMT-1710/00643169568419;MMT-1750/00763000253189-00763000505165.All serial numbers.	234,360 Units Nationwide including C	October 2024 and prior

For additional information, please visit the [FDA website](#).

CDPH Food and Drug Branch
 MS 7602 ● P.O. Box 997435 ● Sacramento, CA 95899-7435
 (916) 650-6500 ● (916) 650-6650 FAX
 Internet Address: www.cdph.ca.gov

