



[Food and Drug Branch – Device Recalls](#)

**California Device Recall Information Sheet**

**Mercury Enterprises, Inc. dba Mercury Medical Recalls Neo-Tee T-Piece Resuscitator**

Recall Date	Product Description	Recalling Firm	Recall Reason
10/9/2024	<b>Neo-Tee T-Piece Resuscitator</b> Neo-Tee T-Piece Resuscitator. Part Numbers 1050805, 1050832. For pediatric emergency respiratory support.	<b>MERCURY ENTERPRISES, INC. DBA MERCURY MEDICAL</b> Clearwater, FL	Manual resuscitator circuit flow controller may come apart due to a missing manufacturing step.

Recall Class	Product Identification	Distribution	Affected Dates
I	<b>Neo-Tee T-Piece Resuscitator</b> Part Number: 1050805; UDI/DI: 10641043508053, 30641043508057; Lot Numbers: 2417650805, 2417750805. Part Number: 1050832; UDI/DI: 10641043508329, 30641043508323; Lot Numbers: 2423550832, 2417850832.	1,300 Units Nationwide	October 2024 and Prior

For additional information, please visit the [FDA website](#).

CDPH Food and Drug Branch  
 MS 7602 ● P.O. Box 997435 ● Sacramento, CA 95899-7435  
 (916) 650-6500 ● (916) 650-6650 FAX  
 Internet Address: [www.cdph.ca.gov](http://www.cdph.ca.gov)

