

## California Device Recall Information Sheet

[Food and Drug Branch – Device Recalls](#)

MICROCARE MEDICAL Recalls Good Vibrations Ultrasonic Cleaner Solution

| Recall Date | Product Description  | Recalling Firm                            | Recall Reason   |
|-------------|--|---|---|
| 10/02/24    | Good Vibrations Ultrasonic Cleaner Solution<br>JorVet J0382G Ultrasonic Instrument Cleaning Solution,<br>SpecClean Ultrasonic Cleaner Solution | MICROCARE<br>MEDICAL<br>Commerce City, Co | Contaminated equipment and packaging caused <i>Klebsiella oxytoca</i> , <i>Enterobacter asburiae</i> , <i>Enterobacter bugandensis</i> , pathogenic and opportunistic bacteria, to occur in ultrasonic detergent, which caused product bloating/leaking, and if used during medical/dental procedures could result in urinary tract, respiratory, bloodstream infections; workers handling contaminated product/instruments also at risk. |

| <b>Recall Class</b> | <b>Product Identification</b>  | <b>Distribution</b>    | <b>Affected Dates</b>  |
|---------------------|--|------------------------|------------------------|
| I                   | Good Vibrations Ultrasonic Cleaner Solution<br>GVUS24-1 Barcode<br>+H816GVUS128H, GVUS128-1<br>+H816GVUSU24% | 66 Units in California | October 2024 and prior |

For additional information, please visit the FDA Website.