

California Device Recall Information Sheet

Food and Drug Branch – Device Recalls

Mirion Technologies (Capintec), Inc. Recalls Captus 4000e Thyroid Uptake System

Recall Date	Product Description	Recalling Firm	Recall Reason
5/24/24	Captus 4000e Thyroid Uptake System. Intended to be used to perform Thyroid Uptakes, Staff Bioassay, and wipe tests.	Healgen Scientific Llc	Complaints of unexpected detachment of the collimator have been reported.

Recall Class	Product Identification	Distribution	Affected Dates
II	Lot Code: Model No: 5430-30151; UDI 859942006096, S/N range 940001 through 941398.	88 Units in California	May 2024 and prior.

For additional information, please visit the [FDA Website](#).