



Food and Drug Branch – Device Recalls

California Device Recall Information Sheet

Percussionaire Corporation Recalls Ipv Therapy In-Line Valve

Recall Date	Product Description	Recalling Firm	Recall Reason
9/27/2024	IPV (Intrapulmonary Percussive Ventilation) Therapy In-Line Valve REF: P5-TEE-20, Qty: 20, Rx Only. IPV (Intrapulmonary Percussive Ventilation) therapy In-Line Valv	PERCUSSIONAIRE CORPORATION Sandpoint, Idaho	Updated labeling to ensure that blue cap on in-line valve is removed, prior to delivery of (Intrapulmonary percussive ventilation) IPV therapy.
9/27/2024	IPV (Intrapulmonary Percussive Ventilation) Therapy In-Line Valve REF: P5-TEE, Qty: 1, Rx Only. IPV (Intrapulmonary Percussive Ventilation) therapy In-Line Valve	PERCUSSIONAIRE CORPORATION Sandpoint, Idaho	Updated labeling to ensure that blue cap on in-line valve is removed, prior to delivery of (Intrapulmonary percussive ventilation) IPV therapy.

Recall Class	Product Identification	Distribution	Affected Dates
I	IPV (Intrapulmonary Percussive Ventilation) Therapy In-Line Valve No UDI/Lots: 240620 240610 240418 240326 230612 WO07066 WO07010 WO06020 WO05019 WO04884 WO04827 WO04756 WO04714 WO04656 WO04625 WO04596 WO04553 WO04398 WO04356 WO04061 WO03846 WO03748 WO03688	2523 Units Nationwide Including California	September 2024 and prior.

I	<p>IPV (Intrapulmonary Percussive Ventilation) Therapy In-Line Valve No UDI/ Lots: 240620, 240610, 240418, 240326, 230612, WO07066, WO07010, WO06020, WO05019, WO04884, WO04827, WO04756, WO04714, WO04656, WO04625, WO04596, WO04553, WO04398, WO04356, WO04061, WO03846, WO03748, WO03688</p>	50,460 Units Nationwide Including California	September 2024 and prior.
---	---	--	---------------------------

For additional information, please visit the [FDA website for IPV Valve](#) and [IPV Valve \(P5-TEE\)](#).

CDPH Food and Drug Branch
 MS 7602 ● P.O. Box 997435 ● Sacramento, CA 95899-7435
 (916) 650-6500 ● (916) 650-6650 FAX
 Internet Address: www.cdph.ca.gov

