



Food and Drug Branch – Device Recalls

California Device Recall Information Sheet

Zoll Medical Corporation Recalls Zoll 731 Ventilator Operator Guide

| Recall Date | Product Description | Recalling Firm | Recall Reason |
|-------------|--|---|---|
| 6/26/2024 | ZOLL 731 Ventilator (EMV+, AEV, Eagle II) - for MRI Compatible Devices - Operator's Guide and Quick Reference Guide (QRGs) Part Numbers: 1. 799-EGL2-02 2. 799-EGL2-02-01 3. 799-EGL2-02-04 ... | ZOLL Medical Corporation Chelmsford, Massachusetts | Operator's Guide & Quick Reference Guide (QRG) ZOLL 731 Ventilator for MRI Compatible Devices. Patient Safety Ventilator MRI information omitted from the manuals. Update to the instructions ensures safe distance from the MRI to ensure proper function of the ventilator and reduce the risk of delay in ventilator therapy |

| Recall Class | Product Identification | Distribution | Affected Dates |
|--------------|--|--------------------------|----------------------|
| I | Product 1 Part Numbers/UDI Zoll 731 Defibrillators: 1. 799-EGL2-02 00847946020910 2. 799-EGL2-02-01 00847946023454 3. 799-EGL2-02-04 00847946036867 ... | 2361 Units Nationwide | June, 2024 and prior |

For additional information, please visit the [FDA website](https://www.fda.gov)

